



# CERTIFICATE OF ACCREDITATION

## The ANSI National Accreditation Board

Hereby attests that

**HEAD acoustics, Inc.**  
**6964 Kensington Road**  
**Brighton, MI 48116**

Fulfils the requirements of

**ISO/IEC 17025:2017**

In the fields of

**CALIBRATION & TESTING**

This certificate is valid only when accompanied by a current scope of accreditation document.

The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).



Jason Stine, Vice President

Expiry Date: 31 May 2026

Certificate Number: L2288



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

## SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

### HEAD acoustics, Inc.

6964 Kensington Road  
Brighton, MI 48116  
Norman Carter 248-486-0099

### CALIBRATION AND TESTING

Valid to: **May 31, 2026**

Certificate Number: **L2288**

### CALIBRATION

#### Acoustics and Vibration

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Filters, Analyzer and Measuring Amplifiers and Measuring Equipment	106 ±2 dB (100 to 500) Hz	0.24 dB	Reference Binaural Head (Headphone Tests Only)
	114 dB 250 Hz, 20 µPa	0.19 dB	
	124 dB 250 Hz, 20 µPa	0.3 dB	
Filters, Analyzer and Measuring Amplifiers and Measuring Equipment Sound Pressure Level	114 dB 20 Hz to 20 kHz	0.98 dB	3PASSflex Background Noise

#### Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
DC Current – Measure	(1 to 10) mA	0.12 mA	6 ½ Digit Multimeter
DC Voltage – Measure	(1 to 10) V (10 to 100) V (100 to 200) V	740 µV 1.3 mV 0.11 V	6 ½ Digit Multimeter
AC Voltage – Measure 20 Hz to 20 kHz	(0.01 to 1) V (1 to 10) V	0.9 mV 5.7 mV	6 ½ Digit Multimeter
AC Voltage Level, Measure Reference Level (0 dB) 1 V <sub>rms</sub> @1 kHz	100 Hz to 20 kHz (-40 to 0) dB (0 to 20) dB	0.009 dB 0.03 dB	6 ½ Digit Multimeter, PEQ V Equalizer

## Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
AC Voltage Level, Generate Reference Level (0 dB) 1 V <sub>rms</sub> @1 kHz	100 Hz to 20 kHz (-40 to 0) dB (0 to 20) dB	0.009 dB 0.03 dB	Function Generator

## Time and Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Frequency – Generate and Measure	20 Hz to 300 kHz	0.02 % of reading	HM8123X Frequency Counter, Function Generator

## TESTING


### Acoustics

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
Sound Pressure Level	ETSI TS 103224	Filters, Analyzer and Measuring Amplifiers and Measuring Equipment	3PASSflex Background Noise

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for all parameters, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope
2. This scope is formatted as part of a single document including Certificate of Accreditation No. L2288.



Jason Stine, Vice President